



Clinical Image

Accidental Breakage of an Aspiration Needle During EBUS-TBNA[☆]

Rotura accidental de aguja de punción transbronquial durante la realización de ecobroncoscopia



Blanca De Vega Sanchez,^{a,*} Sofia Jaurrieta Largo,^b Carlos Disdier Vicente^{a,c}

^a Unidad de Broncoscopias y Técnicas Pleurales, Servicio de Neumología, Hospital Clínico Universitario de Valladolid, Valladolid, Spain

^b Servicio de Neumología, Hospital Clínico Universitario de Valladolid, Valladolid, Spain

^c Centro de Investigación en Red Enfermedades Respiratorias (CIBERES), Spain




Fig. 1. Accidental transbronchial needle breakage during endobronchial ultrasound.

We report the case of an 80-year-old woman with Stage IV pulmonary adenocarcinoma with tumor progression after first-line treatment, in whom a second biopsy was requested. Computed tomography revealed the presence of lymphadenopathies in several sites, including the right paratracheal (Fig. 1A) and subcarinal (Fig. 1B) territories, requiring endobronchial ultrasound. With the appropriate ultrasound visualization, histological samples were obtained from the right paratracheal site with a Cook ECHO-HD-22-EBUS-P® needle, without complications; however, when the first puncture was performed in the subcarinal location, we lost ultrasound vision of the needle and experienced significant resistance on withdrawal. Conventional bronchoscopy was performed immediately, during which a flash of metal was glimpsed in the lateral wall of the carina (Fig. 1C). Needle breakage was suspected, and the absence of the needle tip and needle breakage were confirmed on examination. Using 2 mm Radial Jaw™ 4 (Boston Scientific®) endoscopic forceps, the embedded 15 mm fragment was removed without complications (Fig. 1D).

Endobronchial ultrasound is a safe technique, with few complications, and breakage of the needle during the procedure is exceptional (0.2%).¹ The removal of the fragment is essential in order to prevent long-term complications, currently unknown given the low incidence.²

References

- Asano F, Aoe M, Ohsaki Y, Okada Y, Sasada S, Sato S, et al. Complications associated with endobronchial ultrasound-guided transbronchial needle aspiration: a nationwide survey by the Japan Society for Respiratory Endoscopy. *Respir Res.* 2013;14:50.
- Vial M, O'Connell JO, Grosu HB, Ost DE, Eapen GA, Jimenez CA. Needle fracture during endobronchial ultrasound-guided transbronchial needle aspiration. *Am J Respir Crit Care Med.* 2016;193:213–4.

☆ Please cite this article as: de Vega Sánchez B, Largo SJ, Vicente CD. Rotura accidental de aguja de punción transbronquial durante la realización de ecobroncoscopia. *Arch Bronconeumol.* 2019;55:267.

* Corresponding author.

E-mail address: blancadevegasanchez@gmail.com (B. De Vega Sanchez).